

Conductivity Standard Solutions

Produced and calibrated acc. to the instructions given in ISO Guide 34 and ISO 17025

Traceability to SI of NIST (SRM)

Certificate of analysis prepared acc. to ISO Guide 31:2000 „Reference materials - Contents of certificates and labels“.

DESCRIPTION	VALIDITY <i>months</i>	REF	VOLUME	UNIT
Conductivity water (nominal 0 - 2 $\mu\text{S}/\text{cm}$) traceable to NIST	6	CS0M0S	250	ml
Conductivity Standard 5 $\mu\text{S}/\text{cm}$ at 25°C in 30% n-propanol	3	CS5M0S	500	ml
Conductivity Standard 10 $\mu\text{S}/\text{cm}$ at 25°C in 30% n-propanol	3	CS10M0S	500	ml
Conductivity Standard 15 $\mu\text{S}/\text{cm}$ at 25°C in 30% n-propanol	3	CS15M0S	500	ml
Conductivity Standard 20 $\mu\text{S}/\text{cm}$ at 25°C in 30% n-propanol	3	CS20M0S	500	ml
Conductivity Standard 25 $\mu\text{S}/\text{cm}$ at 25°C in 30% n-propanol	12	CS25M0S	500	ml
Conductivity Standard 50 $\mu\text{S}/\text{cm}$ at 25°C	6	CS50M0S	500	ml
Conductivity Standard 84 $\mu\text{S}/\text{cm}$ at 25°C	6	CS84M0S	500	ml
Conductivity Standard 100 $\mu\text{S}/\text{cm}$ at 25°C	6	CS1P2S	500	ml
Conductivity Standard 111.3 $\mu\text{S}/\text{cm}$ at 25°C	12	CS111M3MS	500	ml
Conductivity Standard 147 $\mu\text{S}/\text{cm}$ at 25°C	6	CS147M0S	500	ml
Conductivity Standard 200 $\mu\text{S}/\text{cm}$ at 25°C	6	CS2P2S	500	ml
Conductivity Standard 500 $\mu\text{S}/\text{cm}$ at 25°C	12	CS5P2S	500	ml
Conductivity Standard 718 $\mu\text{S}/\text{cm}$ at 25°C	12	CS718M0S	500	ml
Conductivity Standard 1000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS1P3S	500	ml
Conductivity Standard 1413 $\mu\text{S}/\text{cm}$ at 25°C	12	CS1413M0S	500	ml
Conductivity Standard 1500 $\mu\text{S}/\text{cm}$ at 25°C	12	CS15P2S	500	ml
Conductivity Standard 2000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS2P3S	500	ml
Conductivity Standard 3000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS3P3S	500	ml
Conductivity Standard 5000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS5P3S	500	ml
Conductivity Standard 10000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS1P4S	500	ml
Conductivity Standard 12880 $\mu\text{S}/\text{cm}$ at 25°C	12	CS1288P1S	500	ml
Conductivity Standard 15000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS15P3S	500	ml
Conductivity Standard 20000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS2P4S	500	ml
Conductivity Standard 30000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS3P4S	500	ml
Conductivity Standard 50000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS5P5S	500	ml
Conductivity Standard 100000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS1P5S	500	ml
Conductivity Standard 150000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS15P5S	500	ml
Conductivity Standard 200000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS2P5S	500	ml
Conductivity Standard 300000 $\mu\text{S}/\text{cm}$ at 25°C	12	CS3P5S	500	ml

